

Thursday, October 28, 2021

Certificate of Analysis

Product Name: Pre-Filled Neutral Buffered Formalin, 480ml**Description:** 10% Neutral Buffered Formalin, Prefilled Container, 480ml, Half filled, 24/cs**Product No.:** PFNBF-480**Lot No.:** 26202**Date of Manufacture:** 2021-10-26**Expiration Date:** 2023-10-01**Materials:**

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Chemical:	NBF-275F	1J010	2023-10-01

QC Testing:

LOT NUMBER: 1J010			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.03 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	9.70% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America**Latex Free:** This product contains no latex

Created By:

Fernando Fana
Quality Assurance Associate