



**AZER SCIENTIFIC™**

701 Hemlock Road, Morgantown, PA 19543  
Phone: 610.524.5810 Fax: 610.901.3046  
www.AzerSci.com

11/01/2021

## CERTIFICATE OF ANALYSIS

**Product Name:** 20mL Pre-Fill Formalin, Prostate Biopsy Kit

**Description:** Kit, 12 place Prostate Biopsy, 12x20mL vials half filled with 10% formalin per kit, 25 kits per case

**Product No.:** 10-820

**Lot No.:** 26701

**Date of Manufacture:** 2021-10-29

**Expiration Date:** 2023-10-01

**Materials:**

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
<b>Chemical:</b>	NBF-275G	1J006	2023-10-01

**QC Testing:**

**Formalin:**

### **LOT NUMBER: 1J006**

PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.06 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

Created By:  
James Osborn  
QA Associate