



AZER SCIENTIFIC™

701 Hemlock Road, Morgantown, PA 19543
Phone: 610.524.5810 Fax: 610.901.3046
www.AzerSci.com

11/3/2021

CERTIFICATE OF ANALYSIS

Product Name: 20mL Pre-Fill Formalin

Description: Prefilled 10% Neutral Buffered Formalin, 20mL, half-filled, Formalock closure 32/pack, 4 packs/case

Product No.: PFNBF-20

Lot No.: 26732

Date of Manufacture: 2021-11-01

Expiration Date: 2023-10-01

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Chemical:	NBF-275G	1J006	2023-10-01
		1J012	2023-10-01

QC Testing:

Formalin:

LOT NUMBER: 1J006			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.06 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS
LOT NUMBER: 1J012			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	6.98 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
James Osborn
QA Associate