



701 Hemlock Road,
Phone: 610.524.5810
info@AzerSci.com

Morgantown, PA 19543
Fax: 610.901.3046
www.AzerSci.com

11/04/2021

CERTIFICATE OF ANALYSIS

Product Name: 10% Neutral Buffered Formalin, 480mL half filled

Description: 10% Neutral Buffered Formalin, 480ml prefills, 24/case

Product No.: PFNBF-480

Lot No.: 26506

Date of Manufacture: 11/03/2021

Expiration Date: 2023-11-01

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Chemical:	NBF-275G	1K002	2023-11-01

QC Testing:

LOT NUMBER: 1K002			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	6.90-7.10 @ 25°C	7.02 @ 25°C	PASS
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)	PASS
Appearance	Clear with No Particulate	Clear	PASS
Labeling	Correct Lot Number and Exp. Date	Correct	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
James Osborn
QA Associate