

Certificate of Analysis

Product No.: 10-831

Description: Kit, 8 place Prostate Biopsy, 8x20mL vials half filled with 10% formalin per kit, 25 kits per case

Lot No.: 26818

Date of Manufacture: 2021-11-10

Expiration Date: 2023-11-01

Chemical:	<u>Part No.</u> RMC-FORMALIN	<u>Lot No.</u> 1K003	<u>Exp. Date</u> 2023-11-01
------------------	---------------------------------	-------------------------	--------------------------------

QC Testing:

LOT NUMBER: 1K003		
PROPERTY	SPECIFICATION	RESULTS
pH	6.90-7.10 @ 25°C	7.05 @ 25°C
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31% (v/v)
Appearance	Clear with No Particulate	Clear

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
James Osborn
Quality Associate