



701 Hemlock Road  
Morgantown PA 19543  
Phone: 610.524.5810 Fax: 610.901.3016  
info@AzerSci.com www.AzerSci.com

# Certificate of Analysis

12/17/2021

**Product No.:** ES613-32

**Description:** Reagent Grade Water, 32oz, 12/case

**Lot No.:** 27247

**Date of Manufacturing:** 12/16/2021

**Expiration Date:** 6/16/2023

## RMC-DIH2O Lot #: 27247

PROPERTY	TEST SPECIFICATION	RESULTS
Appearance	Clear, colorless, odorless	Conforms
Resistivity	10 MΩ /cm	18.12MΩ/cm
Silicates (soluble)	<50 µg/L	3.2 µg/L
Sodium ion	<10.0 µg/L	0.13 µg/L
Chloride ion	<10.0 µg/L	0.16 µg/L
TOC	<50 µg/L	14 µg/L
Bacterial Count	< 10 CFU/ml	0 CFU/ml

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

**Created By:**

James E Osborn  
Quality Associate