

Certificate of Analysis

Product Name: Azer Sterile Saline

Description: Prefilled Buffered saline, 3 ml fill in 10 ml vial, sterile, 100/case

Product No.: ES45044

Lot. No.: 27368

Date of Manufacture: 2021-12-20

Expiration Date: 2022-12-14

Product Specifications:

Ingredients: Sodium Chloride, USP Grade
Disodium Phosphate, Food Grade
Monosodium Phosphate, Food Grade
Deionized Water, NCCLS Type I Reagent Grade

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization before filling process begins.
Packaged in 10 mL vials (3 mL fill)

Sterilization: Electron Beam Sterilization (19.50 - 32.30 kGy Surface Dose)

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container/Cap:	4004-1547	200849	N/A
Chemical:	RMC-ES1243	1L27285	2023-06-14

QC Testing:

LOT NUMBER: 27368			
ATTRIBUTE	SPECIFICATION	RESULTS	PASS / FAIL
Appearance	Clear, colorless, no observable particulate	Pass	PASS
pH	7.0-7.20	7.01	PASS
Assay	0.85% to 0.90%	0.88%	PASS
Sterilization	19.50 - 32.30 kGy Surface Dose	27.40	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:

James Osborn
Quality Associate