

## Certificate of Analysis

**Product Name: Azer Sterile Saline**

**Description: Prefilled saline, 3 ml fill in 10 ml vial, sterile, 100/case**

**Product No.: ES45045**

**Lot. No.: 27606**

**Date of Manufacture: 2022-01-14**

**Expiration Date: 2022-12-27**

**Product Specifications:**

Ingredients: Sodium Chloride, USP Grade  
Deionized Water, NCCLS Type I Reagent Grade

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization before filling process begins.  
Packaged in 10 mL vials (3 mL fill)

Sterilization: Electron Beam Sterilization (19.50 - 32.30 kGy Surface Dose)

**Materials:**

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
<b>Container/Cap:</b>	4004-1547	200849, 200848, 200847	N/A
<b>Chemical:</b>	ES124420L ES1244-1000	27434 2A27532	2023-06-27 2023-07-10

**QC Testing:**

<b>LOT NUMBER: 27606</b>			
<b>ATTRIBUTE</b>	<b>SPECIFICATION</b>	<b>RESULTS</b>	<b>PASS / FAIL</b>
<b>Appearance</b>	Clear, colorless, no observable particulate	Pass	<b>PASS</b>
<b>pH</b>	4.5 – 7.0	5.86	<b>PASS</b>
<b>Assay</b>	0.85% to 0.90%	0.88%	<b>PASS</b>
<b>Sterilization</b>	19.50 - 32.30 kGy Surface Dose	26.90	<b>PASS</b>

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

Created By:

James Osborn  
Quality Assurance Associate