

05/04/2022

CERTIFICATE OF ANALYSIS

Product Name: Zinc Formalin, 120mL, half-filled

Description: Prefilled Zinc Formalin, 120mL, half-filled 24/pack, 2 packs/case

Product No.: PFZF-120

Lot No.: 28556

Date of Manufacture: 2022-05-04

Expiration Date: 2023-11-17

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Chemical:	CUNZF-5-G	1321	2023-11-17

QC Testing:

LOT NUMBER: 1321			
PROPERTY	SPECIFICATION	RESULTS	PASS / FAIL
pH	5.6-6.2 @ 25°C	6.2 @ 25°C	PASS
Formaldehyde Concentration	9.9-10.5% (v/v)	10.5% (v/v)	PASS
Appearance	Clear, colorless solution	Clear	PASS
Physical Inspection	Checked for correct label copy, lot number, expiration date, fill volume, absence of particulate matter and overall aesthetics.	Pass	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
 James Osborn
 Quality Associate