



701 Hemlock Road  
Morgantown PA 19543  
Phone: 610.524.5810 Fax: 610.901.3016  
info@AzerSci.com www.AzerSci.com

# Certificate of Analysis

1/24/2022

**Product No.:** ES124310L

**Description:** Blood Bank Saline, buffered, 7.0-7.2 pH range, 10L cube with spigot, Each

**Lot No.:** 27677

**Date of Manufacturing:** 1/20/2022

**Expiration Date:** 7/17/2023

## Buffered Saline Lot #: 2A27658

PROPERTY	TEST SPECIFICATION	RESULTS
PH	7.0-7.2 @ 25°C	7.04
Assay	0.85% to 0.90%	0.90%
Appearance	Clear, Colorless, No Particulate	Conforms
TAMC	< 10 CFU/ml	< 1 CFU/ml
TYMC	< 10 CFU/ml	< 1 CFU/ml

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

**Created By:**

James E Osborn  
Quality Associate