



701 Hemlock Road
Morgantown PA 19543
Phone: 610.524.5810 Fax: 610.901.3016
info@AzerSci.com www.AzerSci.com

Certificate of Analysis

2/21/2022

Product No.: ES124310L

Description: Blood Bank Saline, buffered, 7.0-7.2 pH range, 10L cube with spigot, Each

Lot No.: 28034

Date of Manufacturing: 2/16/2022

Expiration Date: 8/14/2023

Buffered Saline Lot #: 2B27973

| PROPERTY | TEST SPECIFICATION | RESULTS |
|------------|----------------------------------|------------|
| PH | 7.0-7.2 @ 25°C | 7 |
| Assay | 0.85% to 0.90% | 0.88% |
| Appearance | Clear, Colorless, No Particulate | Conforms |
| TAMC | < 10 CFU/ml | < 1 CFU/ml |
| TYMC | < 10 CFU/ml | < 1 CFU/ml |

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:

James E Osborn
Quality Associate