



701 Hemlock Road
Morgantown PA 19543
Phone: 610.524.5810 Fax: 610.901.3016
info@AzerSci.com www.AzerSci.com

Certificate of Analysis

3/21/2022

Product No.: ES124320L

Description: Blood Bank Saline, buffered, 7.0-7.2 pH range, 20L cube with spigot, Each

Lot No.: 28430

Date of Manufacturing: 3/17/2022

Expiration Date: 9/14/2023

Buffered Saline Lot #: 2C28364

PROPERTY	TEST SPECIFICATION	RESULTS
PH	7.0-7.2 @ 25°C	7.1
Assay	0.85% to 0.90%	0.88%
Appearance	Clear, Colorless, No Particulate	Conforms
TAMC	< 10 CFU/ml	< 1 CFU/ml
TYMC	< 10 CFU/ml	< 1 CFU/ml

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:

James E Osborn
Quality Associate