

Certificate of Analysis

3/2/2022

701 Hemlock Road Morgantown PA 19543 Phone: 610.524.5810 Fax: 610.901.3016 info@AzerSci.com www.AzerSci.com

Product No.: PFNBF-480

Description: Prefilled 10% Neutral Buffered Formalin, 480mL, half-filled 24/case

Lot No.: 28112

Date of Manufacturing: 3/1/2022

Expiration Date: 2/1/2024

10% Formalin Lot #: 2B012			
PROPERTY	TEST SPECIFICATION	RESULTS	
PH	6.90-7.10 @ 25°C	6.98@ 25°C	
Formaldehyde Concentration	9.5-10.5% (v/v)	10.31%(v/v)	
Appearance	Clear, colorless, no observable particulate	Conforms	

10% Formalin Lot #: 2B015			
PROPERTY	TEST SPECIFICATION	RESULTS	
PH	6.90-7.10 @ 25°C	6.95@ 25°C	
Formaldehyde Concentration	9.5-10.5% (v/v)	10.00%(v/v)	
Appearance	Clear, colorless, no observable particulate	Conforms	

Country of Origin: United States of America **Latex Free:** This product contains no latex

Created By:

James E Osborn Quality Associate