



701 Hemlock Road
Morgantown PA 19543
Phone: 610.524.5810 Fax: 610.901.3016
info@AzerSci.com www.AzerSci.com

Certificate of Analysis

4/11/2022

Product No.: ES124420L

Description: Blood Bank Saline, non-buffered, 6.5-7.0 pH range, 20L cube with spigot, Each

Lot No.: 28672

Date of Manufacturing: 4/7/2022

Expiration Date: 10/6/2023

Standard Blood Bank Saline Lot #: 2D28671

PROPERTY	TEST SPECIFICATION	RESULTS
PH	6.5 - 7.0 @ 25°C	6.62
Assay	0.85% to 0.90%	0.88%
Appearance	Clear, Colorless, No Particulate	Conforms
TAMC	< 10 CFU/ml	< 1 CFU/ml
TYMC	< 10 CFU/ml	< 1 CFU/ml

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:

James E Osborn
QA Associate