



701 Hemlock Road  
Morgantown PA 19543  
Phone: 610.524.5810 Fax: 610.901.3016  
info@AzerSci.com www.AzerSci.com

# Certificate of Analysis

4/18/2022

**Product No.:** ES12444G

**Description:** Blood Bank Saline, non-buffered, 6.5-7.0 pH range, 1 gallon, 4/case

**Lot No.:** 28712

**Date of Manufacturing:** 4/13/2022

**Expiration Date:** 10/11/2023

## Standard Blood Bank Saline Lot #: 2D28708

PROPERTY	TEST SPECIFICATION	RESULTS
PH	6.5 - 7.0 @ 25°C	6.63
Assay	0.85% to 0.90%	0.88%
Appearance	Clear, Colorless, No Particulate	Conforms
TAMC	< 10 CFU/ml	< 1 CFU/ml
TYMC	< 10 CFU/ml	< 1 CFU/ml

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

**Created By:**

James E Osborn  
QA Associate