



701 Hemlock Road
Morgantown PA 19543
Phone: 610.524.5810 Fax: 610.901.3016
info@AzerSci.com www.AzerSci.com

Certificate of Analysis

4/25/2022

Product No.: ES611-5G

Description: High Purity Water, 5 gallon cube with spigot

Lot No.: 28805

Date of Manufacturing: 4/22/2022

Expiration Date: 10/22/2023

RMC-DIH2O Lot #: 28805

PROPERTY	TEST SPECIFICATION	RESULTS
Appearance	Clear, colorless, odorless	Conforms
Resistivity	10 MΩ /cm	18.24MΩ/cm
Silicates (soluble)	<50 µg/L	3.2 µg/L
Sodium ion	<10.0 µg/L	0.13 µg/L
Chloride ion	<10.0 µg/L	0.16 µg/L
TOC	<50 µg/L	14 µg/L
Bacterial Count	< 10 CFU/ml	0 CFU/ml

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:

James E Osborn
QA Associate