

Certificate of Analysis

4/22/2022

701 Hemlock Road Morgantown PA 19543 Phone: 610.524.5810 Fax: 610.901.3016 info@AzerSci.com www.AzerSci.com

Product No.: PFNBF-240

Description: Prefilled 10% Neutral Buffered Formalin, 240mL, half-filled 24/case

Lot No.: 28827

Date of Manufacturing: 4/20/2022

Expiration Date: 4/1/2024

10% Formalin Lot #: 2D004		
PROPERTY	TEST SPECIFICATION	RESULTS
PH	6.90-7.10 @ 25°C	6.96@ 25°C
Formaldehyde Concentration	9.5-10.5% (v/v)	10.46% (v/v)
Appearance	Clear, colorless, no observable particulate	Conforms

Country of Origin: United States of America **Created By:** Latex Free: This product contains no latex

James E Osborn QA Associate