

Certificate of Analysis

Product Name: Blood Bank Saline – Buffered Solution

Description: Isotonic Saline Solution, 0.85% w/v, Buffered

Product No.: ES124320L

Lot. No.: 23098

Date of Manufacture: 12/18/2020

Expiration Date: 06/14/2022

Product Specifications:

Ingredients: Sodium Chloride, USP Grade

Disodium Phosphate, Food Grade Monosodium Phosphate, Food Grade

Deionized Water, NCCLS Type I Reagent Grade

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization before

filling process begins.

Packaged in 20 Liter Cubetainers (Manufactured by Hedwin Corporation.)

Materials:

Part No. Exp. Date Lot No. Container: CDC6378 11/18/2020, 11/19/2020 N/A CAP-1 Gal-202012592617 N/A Cap: **RMS-38** Spigot: COM4401-IW 1078478 N/A 0L23112 06/14/2022 Chemical: ES1243-1000

QC Testing:

LOT NUMBER: 0L23112			
ATTRIBUTE	SPECIFICATION	RESULTS	PASS / FAIL
Appearance	Clear, colorless, no observable particulate	Pass	PASS
pH	7.0-7.2	7.07	PASS
Assay	0.85% to 0.90%	0.88%	PASS
Isotonicity (Macro observation)	No hemolysis	Pass	PASS
Isotonicity (Micro observation)	No RBC aggregates	Pass	PASS
TAMC	< 10 CFU/ml	0 CFU/ml	PASS
TYMC	< 10 CFU/ml	0 CFU/ml	PASS

Country of Origin: United States of America **Latex Free:** This product contains no latex

Created By: James Osborn Quality Associate