

## **Certificate of Analysis**

Product Name: Blood Bank Saline – Buffered Solution

Description: Isotonic Saline Solution, 0.85% w/v, Buffered

Product No.: ES124310L

Lot. No.: 25932

Date of Manufacture: 2021-08-19

Expiration Date: 2023-02-16

## **Product Specifications:**

| Ingredients: | Sodium Chloride, USP Grade                  |
|--------------|---|
| -            | Disodium Phosphate, Food Grade              |
|              | Monosodium Phosphate, Food Grade            |
|              | Deionized Water, NCCLS Type I Reagent Grade |

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization before filling process begins.

Packaged in 20 Liter Cubetainers (Manufactured by Hedwin Corporation.)

## Materials:

|                    | <u>Part No.</u>      | Lot No.      | Exp. Date  |
|--------------------|----------------------|--------------|------------|
| <b>Container</b> : | CDC6378              | 07/12/2021   | N/A        |
| Cap:               | CAP-1 Gal-<br>RMS-38 | 202105608211 | N/A        |
| Spigot:            | COM4401-IW           | 1345821      | N/A        |
| Chemical:          | ES1243-1000          | 1H25908      | 2023-02-16 |

## QC Testing:

| LOT NUMBER: 1H25908 |  |            |             |  |  |  |
|---------------------|--|------------|-------------|--|--|--|
| ATTRIBUTE           | SPECIFICATION                                  | RESULTS    | PASS / FAIL |  |  |  |
| Appearance          | Clear, colorless, no observable<br>particulate | Pass       | PASS        |  |  |  |
| рН                  | 7.0-7.2  | 7.03       | PASS        |  |  |  |
| Assay               | 0.85% to 0.90%                                 | 0.89%      | PASS        |  |  |  |
| ТАМС                | < 10 CFU/ml                                    | < 1 CFU/ml | PASS        |  |  |  |
| ТҮМС                | < 10 CFU/ml                                    | < 1 CFU/ml | PASS        |  |  |  |

**Country of Origin:** United States of America **Latex Free:** This product contains no latex

Created By: James Osborn Quality Associate

610-524-5810