

Certificate of Analysis

Product Name: Blood Bank Saline – Standard Solution

Description: Isotonic Saline Solution, 0.85% w/v, Non-buffered

Product No.: ES124420L

Lot. No.: 22159

Date of Manufacture: 09/30/2020

Expiration Date: 03/28/2022

Product Specifications:

Ingredients:	Sodium Chloride, USP Grade		
	Deionized Water, NCCLS Type I Reagent Grade		

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization before filling process begins. Packaged in 20 Liter Cubetainers (Manufactured by Hedwin Corporation.)

Materials:

	Part No.	Lot No.	Exp. Date
Container:	CDC6378	08/25/2020, 09/15/2020	N/A
Cap:	CAP-1 Gal- RMS-38	202007581182	N/A
Spigot:	COM4401-IW	1078477	N/A
Chemical:	ES1244-1000	0I22123	2022-03-28

QC Testing:

LOT NUMBER: 0122123					
ATTRIBUTE	SPECIFICATION	RESULTS	PASS / FAIL		
Appearance	Clear, colorless, no observable particulate	Pass	PASS		
рН	6.5 - 7.0	6.7	PASS		
Assay	0.85% to 0.90%	0.88%	PASS		
Isotonicity (Macro observation)	No hemolysis	Pass	PASS		
Isotonicity (Micro observation)	No RBC aggregates	Pass	PASS		
ТАМС	< 10 CFU/ml	0 CFU/ml	PASS		
TYMC	< 10 CFU/ml	0 CFU/ml	PASS		

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By: Fernando Fana Quality Associate