

Certificate of Analysis

Product Name: Blood Bank Saline – Standard Solution

Description: Isotonic Saline Solution, 0.85% w/v, Non-buffered

Product No.: ES12444G

Lot. No.: 24766

Date of Manufacture: 05/12/2021

Expiration Date: 11/10/2022

Product Specifications:

Ingredients: Sodium Chloride, USP Grade

Deionized Water, NCCLS Type I Reagent Grade

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization before

filling process begins.

Packaged in 4x1 gallon containers with heat sealed lid.

Materials:

 Part No.
 Lot No.
 Exp. Date

 Container:
 P1G-1Gal
 802002
 N/A

 Lid:
 CAP-1 Gal-RMS-38
 202102599242
 N/A

Chemical: ES1244-1000 1E24747 **2022-11-10**

QC Testing:

LOT NUMBER: 1E24747			
ATTRIBUTE	SPECIFICATION	RESULTS	PASS / FAIL
Appearance	Clear, colorless, no observable particulate	Pass	PASS
pH	6.5 - 7.0	6.79	PASS
Assay	0.85% to 0.90%	0.88%	PASS
Isotonicity (Macro observation)	No hemolysis	Pass	PASS
Isotonicity (Micro observation)	No RBC aggregates	Pass	PASS
TAMC	< 10 CFU/ml	0 CFU/ml	PASS
TYMC	< 10 CFU/ml	0 CFU/ml	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By: James Osborn Quality Associate