

Certificate of Analysis

Product Name: Blood Bank Saline – Standard Solution

Description: Isotonic Saline Solution, 0.85% w/v, Non-buffered

Product No.: ES12444G

Lot. No.: 24955

Date of Manufacture: 05/24/2021

Expiration Date: 11/24/2022

Product Specifications:

Ingredients: Sodium Chloride, USP Grade
Deionized Water, NCCLS Type I Reagent Grade

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization before filling process begins.
Packaged in 4x1 gallon containers with heat sealed lid.

Materials:

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container:	P1G-1Gal	802001	N/A
Lid:	CAP-1 Gal-RMS-38	202105605861	N/A
Chemical:	ES1244-1000	1E24945	2022-11-24

QC Testing:

LOT NUMBER: 1E24945			
ATTRIBUTE	SPECIFICATION	RESULTS	PASS / FAIL
Appearance	Clear, colorless, no observable particulate	Pass	PASS
pH	6.5 - 7.0	6.84	PASS
Assay	0.85% to 0.90%	0.88%	PASS
Isotonicity (Macro observation)	No hemolysis	Pass	PASS
Isotonicity (Micro observation)	No RBC aggregates	Pass	PASS
TAMC	< 10 CFU/ml	0 CFU/ml	PASS
TYMC	< 10 CFU/ml	0 CFU/ml	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:
James Osborn
Quality Associate