

## Certificate of Analysis

**Product Name: Blood Bank Saline – Standard Solution**

**Description: Isotonic Saline Solution, 0.85% w/v, Non-buffered**

**Product No.: ES124420L**

**Lot. No.: 25194**

**Date of Manufacture: 06/16/2021**

**Expiration Date: 12/14/2022**

**Product Specifications:**

Ingredients: Sodium Chloride, USP Grade  
Deionized Water, NCCLS Type I Reagent Grade

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization before filling process begins.

Packaged in 20 Liter Cubetainers (Manufactured by Hedwin Corporation.)

**Materials:**

	<u>Part No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>
Container:	CDC6378	05/26/2021, 05/27/2021	N/A
Cap:	CAP-1 Gal- RMS-38	202105605861	N/A
Spigot:	COM4401-IW	1345815	N/A
Chemical:	ES1244-1000	1F25195	2022-12-14

**QC Testing:**

<b>LOT NUMBER: 1F25195</b>			
<b>ATTRIBUTE</b>	<b>SPECIFICATION</b>	<b>RESULTS</b>	<b>PASS / FAIL</b>
Appearance	Clear, colorless, no observable particulate	Pass	<b>PASS</b>
pH	6.5 - 7.0	6.90	<b>PASS</b>
Assay	0.85% to 0.90%	0.88%	<b>PASS</b>
Isotonicity (Macro observation)	No hemolysis	Pass	<b>PASS</b>
Isotonicity (Micro observation)	No RBC aggregates	Pass	<b>PASS</b>
TAMC	< 10 CFU/ml	0 CFU/ml	<b>PASS</b>
TYMC	< 10 CFU/ml	0 CFU/ml	<b>PASS</b>

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

Created By:  
James Osborn  
Quality Associate