

Certificate of Analysis

Product Name: Blood Bank Saline - Standard Solution

Description: Isotonic Saline Solution, 0.85% w/v, Non-buffered

Product No.: ES12444G

Lot. No.: 25748

Date of Manufacture: 2021-08-11

Expiration Date: 2023-02-09

Product Specifications:

Ingredients: Sodium Chloride, USP Grade

Deionized Water, NCCLS Type I Reagent Grade

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization

before filling process begins.

Packaged in 10 Liter Cubetainers (Manufactured by Hedwin Corporation.)

Materials:

 Container:
 P1G-1Gal P1G-1 Gal-RMS-38
 Lot No. B61743
 Exp. Date N/A

 Cap:
 CAP-1 Gal-RMS-38
 202105608211
 N/A

Chemical: ES1244-1000 1H25789 2023-02-09

QC Testing:

LOT NUMBER: 1H25789			
ATTRIBUTE	SPECIFICATION	RESULTS	PASS / FAIL
Appearance	Clear, colorless, no observable particulate	Pass	PASS
pH	6.5 - 7.0	6.55	PASS
Assay	0.85% to 0.90%	0.88%	PASS
TAMC	< 10 CFU/ml	< 1 CFU/ml	PASS
TYMC	< 10 CFU/ml	< 1 CFU/ml	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By: James Osborn QA Associate