

Certificate of Analysis

Product Name: Blood Bank Saline - Standard Solution

Description: Isotonic Saline Solution, 0.85% w/v, Non-buffered

Product No.: ES124410L

Lot. No.: 26016

Date of Manufacture: 08/25/2021

Expiration Date: 02/23/2023

Product Specifications:

Ingredients: Sodium Chloride, USP Grade

Deionized Water, NCCLS Type I Reagent Grade

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization

before filling process begins.

Packaged in 10 Liter Cubetainers (Manufactured by Hedwin Corporation.)

Materials:

Part No. Lot No. Exp. Date Container: CDC5894 07/12/2021 N/A CAP-1 Gal-202105608211 N/A Cap: **RMS-38** Spigot: COM4401-IW 1345821 N/A

Chemical: ES1244-1000 1H25987 2023-02-23

QC Testing:

LOT NUMBER: 1H25987			
ATTRIBUTE	SPECIFICATION	RESULTS	PASS / FAIL
Appearance	Clear, colorless, no observable particulate	Pass	PASS
рН	6.5 - 7.0	6.72	PASS
Assay	0.85% to 0.90%	0.88%	PASS
TAMC	< 10 CFU/ml	0 CFU/ml	PASS
TYMC	< 10 CFU/ml	0 CFU/ml	PASS

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By: James Osborn Quality Associate